

Laboratory Name: M.M.D. KANTAWALA LLP CALIBRATION LABORATORY

A/17, MOHAN ESTATE, KHOKHRA, AHMEDABAD – 380008 GUJARAT, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC – 2346

Page No 4 of 4

Validity 07/10/2024 to 17/08/2026

Last Amended on -----

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
III. <u>Volume</u>^{\$}				
34.	Volume Glassware (Pipette, Burette, Measuring Cylinder, Volumetric Flask)	1 ml to 100 ml	27 μ l	Using Electronic Balance (readability of 0.01 mg & readability of 0.1 mg) and distilled water calibration of Glassware based on Gravimetric Method as per ISO 4787: 2021
		100 ml to 1000 ml	50 μ l	
35.	Micro Pipettes	100 μ l to 1000 μ l	2.6 μ l	Using Electronic Balance (readability of 0.01 mg) and distilled water calibration of Micro Pipette based on Gravimetric Method as per ISO 8655 – 6: 2022