

Laboratory Name: M.M.D. KANTAWALA LLP CALIBRATION LABORATORY

A/17, MOHAN ESTATE, KHOKHRA, AHMEDABAD – 380008 GUJARAT, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC – 2346

Page No 3 of 4

Validity 07/10/2024 to 17/08/2026

Last Amended on -----

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ( $\pm$ )	Remarks
<b>II. <u>Weighing Scale and Balance</u><sup>#</sup></b>				
25.	Mass Comparator and Weighing Balance d $\geq$ 0.0001 mg	0 to 6.1 g	0.0054mg	Using E1 Class Weights by Comparison Method as per OIML R-76 – 1
26.	Mass Comparator and Weighing Balance d $\geq$ 0.001 mg	> 6.1 g to 111 g	0.022 mg	Using E1 Class Weights by Comparison Method as per OIML R-76 – 1
27.	Mass Comparator and Weighing Balance d $\geq$ 0.01 mg	> 111 g to 610 g	0.086 mg	Using E1 Class Weights by Comparison Method as per OIML R-76 – 1
28.	Mass Comparator and Weighing Balance d $\geq$ 0.1 mg	> 610 g to 2300 g	0.4 mg	Using E1 Class Weights by Comparison Method as per OIML R-76 – 1
29.	Weighing Balance d $\geq$ 0.001 g	>2.3 kg to 10.1 kg	4 mg	Using E1 Class Weights by Comparison Method as per OIML R-76 – 1
30.	Weighing Balance d $\geq$ 0.005 g	>10.1 kg to 60 kg	20 mg	Using E1 Class Weights by Comparison Method as per OIML R-76 – 1
31.	Weighing Balance d $\geq$ 1 g	>60 kg to 100 kg	4 g	Using E1, F1 Class Weights by Comparison Method as per OIML R-76 – 1
32.	Weighing Balance d $\geq$ 10 g	>100 kg to 300 kg	10 g	Using F1, M1 Class Weights by Comparison Method as per OIML R-76 – 1
33.	Weighing Balance d $\geq$ 50 g	>300 kg to 500 kg	350 g	Using F1, M1 Class Weights by Comparison Method as per OIML R-76 – 1